



Certificate of Calibration
 Certificate Number: 69579
 Date of Issue: January 22, 2009

"Your Total Test Equipment Solution Provider!"

Calibration Performed For:

JOHN DOE COMPANY, INC.
 1234 MAIN ST
 ANYTOWN, USA 77777
 Phone: 972-999-1111
 Contact: JIMMY GARFUNKLE

Equipment Information

Asset Number: JDC1122
 Manufacturer: AGILENT
 Model Number: E5515C
 Description: WIRELESS COMMUNICATIONS TEST SET
 Serial Number: GQ78796588

Calibration Statistics

Cal Date: 05/20/2008
 Cal Due: 05/20/2010
 Calibration Result: **PASS**
 Cal. Interval: 24 MONTHS
 Performed By: JODI RAY
 Calibration Location: MAT LAB
 Comments/Notes:

Received Condition: IN TOLERANCE
 Returned Condition: IN TOLERANCE
 Temperature: 78 F Humidity: 36 %
 Cal Procedure: MATTX00485
 Department:
 Purchase Order:

Measurement Assurance hereby certifies that...

the above described instrument met or exceeded all published specifications at the time of calibration specified above; and has been calibrated using standards whose accuracies are traceable to the National Institute of Standards and Technology (NIST) within the limitations of the Institute's calibration services, or have been derived from accepted values, physical constants, by ratio or self calibration techniques. All calibration activities performed are in compliance with ANSI/NCSS Z-540-1-1994, MIL STD 45662A, and ISO/IEC 17025 when specified, as well as other national/international systems guidelines. The Quality System is ISO 9001-2000 Certified. This report shall not be reproduced, except in full, without the written permission of Measurement Assurance Technology.

All calibrations, unless otherwise noted, are performed using accuracies of less than or equal to one quarter of the specification of the unit under test. When applicable, the measurement uncertainties include a coverage factor of k=2, having a confidence level of 95%.

Standards Used To Calibrate Equipment

Trace Number	Manufacturer	Model	Description	Last Cal.	Cal. Due Date
A1136A	HEWLETT PACKARD	8903B	AUDIO ANALYZER	10/3/08	10/3/09
A1160	AGILENT	53132A	UNIVERSAL COUNTER	4/3/08	4/3/09
A1178	HEWLETT PACKARD	11667A	POWER SPLITTER	10/23/07	10/23/09
A1215A	HEWLETT PACKARD	E4419A	POWER METER	5/6/08	5/6/10
A1234	HEWLETT PACKARD	ECP-E18A	POWER SENSOR	8/7/07	8/7/08
A1293	AGILENT	E4432B	SIGNAL GENERATOR	2/28/08	2/28/09

Quality Approved: 